

(12) INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

(19) World Intellectual Property
Organization
International Bureau



(43) International Publication Date
10 September 2004 (10.09.2004)

PCT

(10) International Publication Number
WO 2004/077086 A1

(51) International Patent Classification⁷: **G01R 33/563**,
33/561

(21) International Application Number:
PCT/IB2004/050131

(22) International Filing Date: 18 February 2004 (18.02.2004)

(25) Filing Language: English

(26) Publication Language: English

(30) Priority Data:
03100506.9 28 February 2003 (28.02.2003) EP

(71) Applicant (for all designated States except US): **KONINKLIJKE PHILIPS ELECTRONICS N.V.** [NL/NL];
Groenewoudseweg 1, NL-5621 BA Eindhoven (NL).

(72) Inventors; and

(75) Inventors/Applicants (for US only): **FUDERER, Miha**
[BE/NL]; c/o Prof. Holstlaan 6, NL-5656 AA Eindhoven

(NL). **HARVEY, Paul, R.** [GB/NL]; Prof. Holstlaan 6,
NL-5656 AA Eindhoven (NL). **VAN DEN BRINK, Johan,**
S. [NL/NL]; c/o Prof. Holstlaan 6, NL-5656 AA Eindhoven
(NL).


(74) Agent: **COHEN, Julius, S.**; Prof. Holstlaan 6, NL-5656
AA Eindhoven (NL).

(81) Designated States (unless otherwise indicated, for every
kind of national protection available): AE, AG, AL, AM,
AT, AU, AZ, BA, BB, BG, BR, BW, BY, BZ, CA, CH, CN,
CO, CR, CU, CZ, DE, DK, DM, DZ, EC, EE, EG, ES, FI,
GB, GD, GE, GH, GM, HR, HU, ID, IL, IN, IS, JP, KE,
KG, KP, KR, KZ, LC, LK, LR, LS, LT, LU, LV, MA, MD,
MG, MK, MN, MW, MX, MZ, NA, NI, NO, NZ, OM, PG,
PH, PL, PT, RO, RU, SC, SD, SE, SG, SK, SL, SY, TJ, TM,
TN, TR, TT, TZ, UA, UG, US, UZ, VC, VN, YU, ZA, ZM,
ZW.

(84) Designated States (unless otherwise indicated, for every
kind of regional protection available): ARIPO (BW, GH,

[Continued on next page]

(54) Title: MOVING TABLE MRI WITH SUBSAMPLING



(57) Abstract: A novel magnetic resonance imaging method is described, for forming an image of an object from a plurality of signals sampled in a restricted homogeneity region of a main magnet field of a magnetic resonance imaging apparatus. A patient disposed on a table is moved continuously through the bore of the main magnet and spins in a predetermined area of the patient are excited by an excitation pulse from a transmitter antenna, such that an image is formed over a region exceeding largely the restricted region. Data is undersampled in the restricted region by means of at least one receiver antenna in a plurality of receive situations being defined as a block of measurements contiguous in time having preserved magnetisation and presaturation conditions within the excited area of the patient. Fold-over artefacts due to said undersampling are unfolded by means of the known sensitivity pattern of the receiver antenna and/or the properties of selected factors determining said receive situations.

WO 2004/077086 A1



GM, KE, LS, MW, MZ, SD, SL, SZ, TZ, UG, ZM, ZW), Eurasian (AM, AZ, BY, KG, KZ, MD, RU, TJ, TM), European (AT, BE, BG, CH, CY, CZ, DE, DK, EE, ES, FI, FR, GB, GR, HU, IE, IT, LU, MC, NL, PT, RO, SE, SI, SK, TR), OAPI (BF, BJ, CF, CG, CI, CM, GA, GN, GQ, GW, ML, MR, NE, SN, TD, TG).

Declaration under Rule 4.17:

- *as to applicant's entitlement to apply for and be granted a patent (Rule 4.17(ii)) for the following designations AE, AG, AL, AM, AT, AU, AZ, BA, BB, BG, BR, BW, BY, BZ, CA, CH, CN, CO, CR, CU, CZ, DE, DK, DM, DZ, EC, EE, EG, ES, FI, GB, GD, GE, GH, GM, HR, HU, ID, IL, IN, IS, JP, KE, KG, KP, KR, KZ, LC, LK, LR, LS, LT, LU, LV, MA, MD, MG, MK, MN, MW, MX, MZ, NA, NI, NO, NZ, OM, PG, PH, PL, PT, RO, RU, SC, SD, SE, SG, SK, SL, SY, TJ,*

TM, TN, TR, TT, TZ, UA, UG, UZ, VC, VN, YU, ZA, ZM, ZW, ARIPO patent (BW, GH, GM, KE, LS, MW, MZ, SD, SL, SZ, TZ, UG, ZM, ZW), Eurasian patent (AM, AZ, BY, KG, KZ, MD, RU, TJ, TM), European patent (AT, BE, BG, CH, CY, CZ, DE, DK, EE, ES, FI, FR, GB, GR, HU, IE, IT, LU, MC, NL, PT, RO, SE, SI, SK, TR), OAPI patent (BF, BJ, CF, CG, CI, CM, GA, GN, GQ, GW, ML, MR, NE, SN, TD, TG)

Published:

- *with international search report*

For two-letter codes and other abbreviations, refer to the "Guidance Notes on Codes and Abbreviations" appearing at the beginning of each regular issue of the PCT Gazette.

10/547196

MOVING TABLE MRI WITH SUBSAMPLING

The invention relates to a magnetic resonance (MR) method for forming an image of an object from a plurality of signals sampled in a restricted homogeneity region of a main magnet field of the magnetic resonance imaging apparatus, whereas a patient disposed on a table is moved continuously through the bore of the main magnet, such that an image is
5 formed over a region exceeding largely the restricted region of a main magnet field of the magnetic resonance imaging apparatus.

The invention also relates to an MR apparatus and a computer program product for carrying out such a method.

Since it is quite uncomfortable for a patient being disposed in a large bore of
10 the main superconductive magnet of an MR imaging apparatus, there is a trend to use shorter magnets while moving the patient table or couch through the bore of the magnet. Such an apparatus is described e.g. in EP-A-1 024 371. In an MR imaging apparatus excitation pulses are applied to a restricted region of the magnet bore, in which the field is uniform. The data samples collected are Fourier transformed to form a volumetric image of the restricted
15 region. A motor continuously moves a patient couch so that a region of interest passes through the region of good field. The collected data samples are corrected to compensate for the motion so that a volumetric image is formed of greater length than that of the restricted region.

The MR imaging as mentioned above takes quite a long time since it is
20 performed with conventional imaging sequences and even an overlap between successively imaged volumes is preferred in order to discard artefacts produced by errors in the central region. Obviously, the r.f. transmit coil and the r.f. receive coil are disposed in the reference frame of the main magnet, i.e. in its direct vicinity. However, imaging by body coils disposed on different places of a patient was not an issue in this reference. Anyway, it is known that
25 using body coils may cause additional artefacts which cannot be suppressed in an easy manner.

Therefore, it is an object of the present invention to improve the MR imaging method with a relative short main magnet and a moving bed in such a manner, that imaging

will be performed in a much shorter time period. Another object of the present invention is to provide an MR apparatus and a computer program product for carrying out the method.

These objects are achieved by a method as claimed in Claim 1, by an MR apparatus as claimed in Claim 11 and by a computer program product as claimed in Claim 13.

The invention has the main advantage, that shorter main magnets with a restricted homogeneity region can be used whereas data can be sampled in a much shorter time due to undersampling and therefore the scan time is largely reduced in relation to conventional MR imaging with a moving table or bed. Another advantage of the present invention is that the direction of the field-of-view can be oriented in any direction, which means that e.g. flow artefacts can be oriented in an appropriate manner or to orient single slices to some preferred orientation with respect to the tissue structure of the patient.

These and other advantages of the invention are disclosed in the dependent claims and in the following description in which an exemplified embodiment of the invention is described with respect to the accompanying drawings. Therein shows:

Fig. 1 diagrammatically a magnetic resonance imaging system in which the present invention is used,

Fig. 2 an example of a acquisition scheme,

Fig. 3 an acquisition scheme in k-space with a moving bed, and

Fig. 4 a simplified presentation for explanation of the present invention.

Basic principles and notations

In Fig. 1 a patient 1 is placed fixedly to a moving table or bed 2. A set of local coils LC_1 to LC_5 is located fixedly to the patient 1, thus the arrangement "patient 1", "bed 2" and "local coils LC_1 to LC_5 " are disposed in a fixed relationship to each other in a coordinate system 3 which is related to the patient. A vector \vec{x} denotes the position relative to this coordinate system or frame of reference 3. Further a main magnet 4 is provided with a bore 5, in which the bed 2 can be moved to and fro in direction of arrow 6. The magnet/gradient system 7 comprises the main magnet 4 and a number of receive coils 8 - generally one or none - and a transmitter coil (not depicted). These receive coils 8 and transmit coils are denoted as being "global". The elements of the magnet/gradient system 7 are all related in another fixed relationship which is defined by another coordinate system 9 related to the

main magnet 4. It is further assumed that the receive coils 8 and the transmit coils are large and have smooth sensitivity patterns in the direction of the bed motion (arrow 6). The magnet coordinate system 9 is offset by a distance $\bar{\Delta}$ relative to the patient coordinate system 3. The property of the patient 1, e.g. the proton density of his body, is obviously given in the patient coordinate system 3, expressed as $\rho(\bar{x})$. Ideally, the same holds for the receive-sensitivity pattern $S_i(\bar{x})$ of one of the local coils LC_i , where $i = 1$ to 5. On the other hand, the sensitivity pattern of the receive coil 7 is written as $g(\bar{x} - \bar{\Delta})$.

Now the expression "receive situation" is introduced, which means a block of measurements contiguous in time, during which the table 2 has travelled a relatively small distance (e.g. one half or one third or something in between) of the "homogeneity size" of the main magnet 4. During a single receive situation all magnetisation preparations (like presaturation slabs, volume selections etc.) "travel" in principle with the patient 1. Also during a single receive situation, all of the interesting k-space is covered, except for any (potentially significant) undersampling. The full scan over k-space covers K receive situations. In principle, the MR acquisition sequence is identical over all receive situations, except that the "magnetisation preparations" may change place and that there may be some offset in the acquired k-space grid. This is however a refinement of the concept.

During any receive situation, every coil LC_1 to LC_5 receives information from the patient 1. And so $(I + J)$ sets of information are delivered, where I is the number of local receive coils and J is the number of global receive coils. Some of these sets may be practically zero, e.g. if a receive coil LC_1 to LC_5 is completely outside the homogeneity volume of main magnet 4. There are K receive situations. Thus, in total, information has been gathered in $K * (I + J)$ "receive instances".

25 Basic idea of the invention

The clinical doctor wants information from a relatively large volume of interest of the patient 1. Data is acquired while the table or bed 2 is moving relative to the main magnet 4. The displacement $\bar{\Delta}$ thereof relative to the magnet 4 is known at any time. The whole of the MR acquisition is split up in a number of receive situations. During each situation, the excitation and presaturation profiles "travel" with the patient 1. The acquired data are phase-corrected for the offset of table 2 at which they are acquired. During a receive situation, information from the patient 1 is acquired using any MR imaging sequence of any orientation. In practice it may be useful either to orient artefacts (e.g. flow artefacts) in an

appropriate direction, or to orient single slices to some preferred orientation with respect to a specific human-tissue structure. However, the information may be seriously undersampled. An example of this method is sketched in Fig. 2. For every receive situation, we acquire, nominally the volume indicated by a box 11. This means that a multitude of points 12 will
5 fold onto the results of this receive situation. This multitude may be a one-dimensional row, a two-dimensional row or even a continuum of folding locations which is the case in e.g. spiral acquisition.

Now, every receive situation has the same subsampling pattern. The only difference between different receive situations is the weighing with respect to the properties
10 of following selected factors:

1. the excitation or saturation profile (with respect to the patient 1),
2. the transmit pattern of the transmit coil,
3. the receive patterns of the fixed receive coils 8, and
4. the frequency-response pattern of the receiver system.

15 Concerning items 2 and 3 it is remarked that in theory, with a moving table 2, these patterns varies also *within* a situation - but if the motion per situation is not exceedingly large, this may be neglected.

In addition, every receive coil (whether local or global) has its own sensitivity pattern. So every receive instance, which is the combination of a receive coil and a receive
20 situation, "sees" a different "overall pattern", which is a product of the coil-sensitivity pattern and the combined pattern (excitation, saturation, transmit and frequency-response) of the receive situation.

The multitude of folding points 12 is unfolded using all the "overall patterns" of every receive instance. This is done in a manner as in the SENSE method (see e.g. K.
25 Pruessmann et. al. in Proc. ISMRM, 1998, abstracts pp. 579, 799, 803 and 2087) - although, formally, the unfolding matrix may be larger. Nevertheless, the system of equations may be reasonably stable: e.g. for each situation, many folding points 12 may practically "see" a zero pattern, because they fall outside the excitation slab. It is further to be noted that the table motion is flexible: it may be linear, non-linear and even non-monotonic or two-dimensional.
30 The restriction is that it must not travel excessively during one receive situation, e.g. no more as half the homogeneity volume.

Measuring the patternsCalibration

In SENSE, a map of coil sensitivity profiles must be available. Such a map can be provided in a reference or coarse calibration measurement with a well-defined phantom.

- 5 Another possibility is using a low-resolution map of the ratio of signals from an array coil element and from a reference coil of the patient on the table. In the present invention a similar reference measurement is provided. Although data can generally be acquired at any table-motion pattern, including two-dimensional table motion, here a linear one-dimensional movement will be used for data acquisition.

10 For a better understanding of the involved principles, first a review:

1. At each position of the patient (\vec{x}) some position-dependent tissue-properties (all data about proton density, T_1 , T_2 , flow, diffusion...) are known
2. Also at each position of the patient (\vec{x}) the sensitivity pattern $S_i(\vec{x})$ of the local coils LC_i can be obtained. Although there might be some dependence on $\vec{\Delta}$ due to coupling to global coils, this is neglected here.
- 15 3. The pattern of all the coils (global or local, transmit or receive) may be influenced by dielectric resonances, particularly at main magnet fields of 3T and higher. It is assumed that this effect can be seen as purely multiplicative on the coil sensitivity of all coils - an effect that is only dependent on the patient and the position therein, so not on the coil or coil-position. For a main magnet field of 1,5T the effect is negligible. The dielectric resonance effect is denoted as $D(\vec{x})$. Each coil actually "sees" with a sensitivity $D(\vec{x}) \cdot S_i(\vec{x})$.
- 20 4. The global coils also have their sensitivity pattern, which is approximated as being independent on the actual position of the patient within the magnet frame (except, of course, for dielectric resonance effect, which is explicitly accounted for). This sensitivity pattern is expressed as $g_j(\vec{x} + \vec{\Delta})$.
- 25 5. The total receive-pattern also may contain effects of receiver frequency response, denoted as $F(\vec{x} + \vec{\Delta})$ at a given readout gradient direction.

For a given measurement type, the transverse magnetisation is a non-linear effect depending on both the transmit pattern - i.e. on the position in the magnet coordinate system 9, if a global coil is used - and the tissue property - i.e. on the position in the patient coordinate system 3. Since this is very difficult to handle, an approximation is necessary, which can be achieved by using an FFE measurement with very small tip angles (about 5 degrees or less). In that case, the transverse magnetisation can be approximated as

30

$f(\vec{x}) \cdot \alpha(\vec{x} + \vec{\Delta})$, where $f(\vec{x})$ is the "spin density" function, which is a tissue property for a given measurement (in the proposed case, it mainly depends on proton density). α is the tip angle, which is approximated as being proportional to $D(\vec{x}) \cdot g_{transmit}(\vec{x} + \vec{\Delta})$.

5 Problems

For the reconstruction of the proposed method as described in more detail below, ideally one would like to know all information of $S_i(\vec{x})$, $g_j(\vec{x} + \vec{\Delta})$ and $D(\vec{x})$ (but the last one is not crucial). The value of $F(\vec{x} + \vec{\Delta})$ is assumed to be known beforehand, and the value of $f(\vec{x})$ is (at least in principle) not an interesting result of the calibration measurement.

The problem is that not all these parameters are measured. For a given table position $\vec{\Delta}$, the following is measured:

- the signal from local coils: $F(\vec{x} + \vec{\Delta}) \cdot S_i(\vec{x}) \cdot D(\vec{x}) \cdot g_{transmit}(\vec{x} + \vec{\Delta}) \cdot D(\vec{x}) \cdot f(\vec{x})$.
- the signal from global coils: $F(\vec{x} + \vec{\Delta}) \cdot g_j(\vec{x} + \vec{\Delta}) \cdot D(\vec{x}) \cdot g_{transmit}(\vec{x} + \vec{\Delta}) \cdot D(\vec{x}) \cdot f(\vec{x})$.

15 A plain division of those two signals results in $S_i(\vec{x}) / g_j(\vec{x} + \vec{\Delta})$ - which is of no direct use if $g_j(\vec{x} + \vec{\Delta})$ is not uniform.

First solution

20 The patterns of the global coils are assumed to be known (i.e. at least one of them). In practice, that function may be load-dependent, but this can be overcome by the fact that there is a calibrated load-dependency of the sensitivity function, and the load can be easily measured.

A number of discrete reference-scan segments are acquired. During such a segment, the table stands still. Each segment results in a full "image". The table is then stepped over a number of segments, so for a number of different values of $\vec{\Delta}$. The knowledge of $g_j(\vec{x} + \vec{\Delta})$ allows calculating the matrix $S_i(\vec{x})$. Mathematically, each segment would give the result over the full extent of \vec{x} , but in practice, each segment will provide accurate results only over a sub-range of all \vec{x} . These partially overlapping ranges can be combined using least-squares fits, resulting in the full map of $S_i(\vec{x})$.

Second solution

The three-dimensional measurement is oriented in such a way that the frequency-encoding is parallel to the table motion direction - the z-direction. The sampling bandwidth is very high in that direction. The other two directions are phase-encodings, one of which is "fast" compared to table motion. The "slow" phase-encoding direction is called k_y . In the patient coordinate system 9 the main magnet 4 is slowly moving with respect to the patient. During the motion of the magnet, profiles are acquired with linearly increasing k_y . In Fig. 3 a line segment 13 indicates the portion of the patient 1 that is detected by the homogeneity volume of the main magnet 4.

10 L is the distance that the main magnet 4 covers for a full set of phase-encoding profiles. The table speed is arranged in such a way that L is maximally half of the homogeneity volume - but less (e.g., 1/4) is preferable. In Fig. 3 L is chosen at 1/2. The consequence is that every profile has been sampled for at least two different positions of the main magnet 4 (except for the outer edge regions, which are disregarded here). For example a
15 profile k_y has been measured for both the displacement $\bar{\Delta}_1$ and for the displacement $\bar{\Delta}_2$. This allows to approximate the profile that would have been measured if the main magnet 4 were at position $\bar{\Delta}_{ref}$ (as long as this value is in between $\bar{\Delta}_1$ and $\bar{\Delta}_2$). This can be done by interpolation (e.g. a linear interpolation if two profiles are available, or a cubic spline if there are four profiles). That can be done for every profile, which allows a complete reconstruction
20 of the sensitivity-calibration measurement at magnet-offset $\bar{\Delta}_{ref}$.

However, this can also be done for any value of $\bar{\Delta}_{ref}$. That principle allows acquiring:

- the signal from local coils: $F(\vec{x} + \vec{\Delta}) \cdot S_l(\vec{x}) \cdot D(\vec{x}) \cdot g_{transmit}(\vec{x} + \vec{\Delta}) \cdot D(\vec{x}) \cdot f(\vec{x})$.
- the signal from global coils: $F(\vec{x} + \vec{\Delta}) \cdot g_j(\vec{x} + \vec{\Delta}) \cdot D(\vec{x}) \cdot g_{transmit}(\vec{x} + \vec{\Delta}) \cdot D(\vec{x}) \cdot f(\vec{x})$.

25 This holds, in contrary to the "problem section, not only for one value of $\bar{\Delta}$ but for all values thereof. This allows to:

- separate the signal from local coils into $F(\vec{x} + \vec{\Delta}) \cdot g_{transmit}(\vec{x} + \vec{\Delta})$ and $S_l(\vec{x}) \cdot D(\vec{x}) \cdot D(\vec{x}) \cdot f(\vec{x})$, and
- separate the signal from global coils into $F(\vec{x} + \vec{\Delta}) \cdot g_j(\vec{x} + \vec{\Delta}) \cdot g_{transmit}(\vec{x} + \vec{\Delta})$ and
30 $D(\vec{x}) \cdot D(\vec{x}) \cdot f(\vec{x})$.

This allows the derivation of $S_j \bar{x}$ and of $g_j(\bar{x} + \bar{\Delta})$. Actually, if F is known, then also $g_{transmit}(\bar{x} + \bar{\Delta})$ could be derived, but unfortunately not the dielectric resonance $D(\bar{x})$ - but that is not crucial for the sequel.

5 Receiving and reconstructing the actual data

In the section "basic idea of the invention" the receiving and reconstruction of data has been explained to a great extent. The data is acquired using relatively large steps between profiles, resulting in a (nominally) small "folding volume", i.e., into lots of folding. For non-cartesian sequences this means that data is sampled relatively sparse. The full extent of k-space is acquired during one scan situation, and the displacement during one scan situation is a fraction (e.g. 1/3) of the homogeneity volume. Every incoming sample is acquired at a very specific value of $\bar{\Delta}_{sample}$. This is taken into account by multiplying every sample of the incoming data with $\exp(-ikT\bar{\Delta}_{sample})$. Here, T is the coordination transformation from patient coordinates to scan coordinates, which is relevant for oblique scanning. That operation "displaces" the acquired data to the centre of the patient, even if that is way out of the homogeneity volume. Obviously, one can simplify things by correcting with a fixed $\bar{\Delta}_{profile}$ per profile.

- The scan situation has been acquired at an average offset $\bar{\Delta}_{ave}$. This is relevant to estimate:
- 20 - the transmit profile for that situation: $g'_{transmit}(\bar{x}) = g_{transmit}(\bar{x} + \bar{\Delta}_{ave})$,
 - similarly for the global-coils receive profile, and
 - similarly for the frequency response function for that situation.

The excitation profiles and presaturation profiles are fixed to the patient coordinate system, i.e. travelling with the table 2 as seen from the main magnet 4. So these are known in \bar{x} . Also the local coil profiles are known in \bar{x} (see calibration section). All these coil profiles can be multiplied into an overall receive pattern for each coil. In total, if there are K receive situations, then information has been gathered in $K \cdot (I + J)$ receive instances. Each instance has a different overall receive pattern. This allows for a SENSE-like reconstruction, if there are less than $K \cdot (I + J)$ folded points over the entire patient 1. In principle, the acquisition stays the same over all receive situations. As a refinement, the phase encoding may be offset by a small amount. This will give some extra phase-encoding,

on top of all other mentioned encodings, which will improve the stability of the SENSE-reconstruction.

Geometric correction

5 All of above reasoning assumes a perfectly linear gradient system. It assumes that a table-displacement of $\bar{\Delta}$ will cause (a) some different weighting, which has been extensively treated above, and (b) a phase modification of $\exp(ik.T\bar{\Delta}_{sample})$ to a sample. Unfortunately, geometric distortion is another significant complicating factor. An image acquired at offset $\bar{\Delta}_1$ is geometrically distorted when compared to an image acquired at offset
 10 $\bar{\Delta}_2$, even if all weightings and the displacement ($\bar{\Delta}_2 - \bar{\Delta}_1$) have been accounted for.

This problem can be solved since the distortions tend to be small where the excitation and transmit profiles tend to be large. This can be accomplished (a) by making sequences in such a way that there is not too much excitation in geometrically inhomogeneous areas, and (b) by designing the system with low transmit-sensitivity in these
 15 areas. The solution is an iterative reconstruction, the explanation thereof is given visually in Fig. 4. The main idea is, in essence, to first reconstruct the resulting image by disregarding the presence of geometric distortion. Of this result, all image regions are distorted in as many ways as there are scan situations. The distorted minus the undistorted versions provide an estimate of "folding distortion-errors". If the contributions (or weights) of every receive-
 20 instance to the final image are known, the estimate of the folding distortion-errors can be extracted from the result.

In Fig. 4 (a) a very simplified patient 15 is depicted which diagrammatically consists of two parts at different positions: a first part shown as block 16 and a second part shown as heart 17. Fig. 4 (b) shows the imaging while the block 16 is largely in the centre of the main magnet and the heart 18 is more remote and thus distorted. As shown in Fig. 4 (c)
 25 the distorted heart 18 folds onto the block 16, due to undersampling. Yet, the heart 18 is fainter, since it has seen less excitation. In another receive-situation according to Fig. 4 (d), the heart 17 is in the centre of the main magnet 4 and the block 19 is more remote and thus distorted. Due to undersampling, this is measured in that receive-situation, as shown in Fig. 4
 30 (e). So the images in Fig. 4 (c) and Fig. 4 (e) are actually measured, Fig. 4 (f) shows the result after SENSE-unfolding, which consists of two linear combinations of Fig. 4 (c) and Fig. 4 (e). It is imperfect, e.g. the right part of it consists of the block 16 plus an unsubtracted heart 18 plus a "wrongly subtracted" heart 17 (dashed) and a second-order error (dotted). In

Fig. 4 (g) is shown a purposely distortion in the same manner as in (b), under the condition that (f) is available. Fig. 4 (h) shows the final result where part of (f) is subtracted from its folded counterpart, and part of (g) is added, resulting in only second-order artefacts.

CLAIMS:

1. A magnetic resonance imaging method for forming an image of an object from a plurality of signals sampled in a restricted homogeneity region of a main magnet field of a magnetic resonance imaging apparatus, wherein a patient disposed on a table is moved continuously through the bore of the main magnet and spins in a predetermined area of the patient are excited by an excitation pulse from a transmitter antenna, such that an image is formed over a region exceeding largely the restricted region, characterized in that data is undersampled in the restricted region by means of at least one receiver antenna in a plurality of receive situations being defined as a block of measurements contiguous in time having preserved magnetisation and presaturation conditions within the excited area of the patient, and fold-over artefacts due to said undersampling are unfolded by means of the sensitivity pattern of the receiver antenna and/or the properties of selected factors determining said receive situations.
2. A method as claimed in Claim 1, characterized in that the fold-over artefacts are unfolded by means of the excitation profile as selected factor.
3. A method as claimed in Claim 1 or 2, characterized in that the fold-over artefacts are unfolded by means of the magnetisation and presaturation profile as selected factor.
4. A method as claimed in one of Claims 1 to 3, characterized in that the fold-over artefacts are unfolded by means of the frequency-response pattern of the receiver as selected factor.
5. A method as claimed in one of Claims 1 to 4, characterized in that during sampling of data within the restricted region the table will be moved over at most one half of the size of the restricted region.

6. A method as claimed in one of Claims 1 to 5, characterized in that at least one global receiver antenna disposed in a fixed relationship to the main magnet system and a plurality of local receiver antennae disposed in a fixed relationship to the patient on the table are provided.

5

7. A method as claimed in one of claims 1 to 6, characterized that a reference scan is provided for obtaining the sensitivity pattern of the receiver antenna.

8. A method as claimed in claim 7, characterized in that data is acquired in a number of discrete reference scan segments at different table positions in the restricted region, whereas during each reference scan the table stands still, and from the data obtained by the reference scans the sensitivity patterns of the local coils are calculated.

9. A method as claimed in claim 7 or 8, characterized in that the table is moved at a speed which is less than half of the homogeneity volume of the main magnet over the scan time of k-space, whereas k-space is continuously scanned in a row-to-row manner, and the sensitivity profiles of the local coils are calculated by interpolation of the measured profiles at different table positions.

10. A method as claimed in any of claims 1 to 9, characterized in that the sampled data is reconstructed in an iterative manner, in that data sampled largely offset of the centre of the main magnet, which is folding-in on the data sampled in the centre of the main magnet, is purposely distorted such that a undistorted image is reconstructed and subtracted from the fold-in image.

25

11. A magnetic resonance imaging apparatus for obtaining an MR image from a plurality of signals comprising:

- a main magnet,
- a transmitter antenna for excitation of spins in a predetermined area of the patient,
- 30 - at least one receiver antenna for sampling a plurality of signals in a restricted homogeneity region of the main magnet field,
- a table for bearing a patient,
- means for continuously moving the table through the bore of the main magnet,

- means for adapting the frequency profile of the transmitter antenna and the frequency profile of the receiver antenna to the position of the continuously moving table relative to a reference point,
- means for sampling the data in the restricted region in a given receive situation being defined as a block of measurements contiguous in time with preserved magnetisation and presaturation conditions in the excited area of the patient, and
- means for unfolding fold-over artefacts due to undersampling by means of the sensitivity pattern of the receiver antenna and/or the properties of selected factors determining said receive situations.

10

12. Apparatus according to Claim 11, further comprising at least one global receiver antenna disposed at a fixed relationship to the main magnet system and a plurality of local receiver antennae disposed in a fixed relationship to the patient on the table.

15

13. A computer program product stored on a computer usable medium for forming an image by means of the magnetic resonance method, comprising a computer readable program means for causing the computer to control the execution:

20

- creating a main magnetic field by a main magnet,
- excitation of spins in a predetermined area of the patient by a transmitter antenna,
- sampling a plurality of signals in a restricted homogeneity region of the main magnet field by at least one receiver antenna,
- continuously moving a table bearing a patient through the bore of the main magnet,
- adapting the frequency profile of the transmitter antenna and the frequency profile of the receiver antenna to the position of the continuously moving table relative to a reference point,
- undersampling the data in the restricted region in a given receive situation being defined as a block of measurements contiguous in time with preserved magnetisation and presaturation conditions in the excited area of the patient, and
- unfolding fold-over artefacts due to said undersampling by means of the sensitivity pattern of the receiver antenna and/or the properties of selected factors determining said receive situations.

25

30


$1/2$ 

FIG.1




FIG.2

2/2




FIG. 3




FIG. 4

INTERNATIONAL SEARCH REPORT

International Application No
PCT/IB2004/050131

A. CLASSIFICATION OF SUBJECT MATTER
IPC 7 G01R33/563 G01R33/561

According to International Patent Classification (IPC) or to both national classification and IPC

B. FIELDS SEARCHED

Minimum documentation searched (classification system followed by classification symbols)
IPC 7 G01R

Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched

Electronic data base consulted during the international search (name of data base and, where practical, search terms used)

EPO-Internal, INSPEC, PAJ, WPI Data, BIOSIS, EMBASE

C. DOCUMENTS CONSIDERED TO BE RELEVANT

Category *	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
X	US 2002/021128 A1 (KUHARA SHIGEHIDE) 21 February 2002 (2002-02-21) paragraphs '0113!-'0165!; figures 10-18	1-8, 11-13
X	US 2003/004408 A1 (ZHU YUDONG) 2 January 2003 (2003-01-02) paragraphs '0018!-'0049!; figures 2-5,7	1,2,5,7, 11,13
E	EP 1 398 642 A (GEN ELECTRIC) 17 March 2004 (2004-03-17) paragraphs '0018!-'0040!; figures 2-4	1,2,5, 11,13
A	EP 1 205 760 A (UNIV LELAND STANFORD JUNIOR) 15 May 2002 (2002-05-15) paragraphs '0017!-'0027!; figure 3 -/-	1,11,13

☒ Further documents are listed in the continuation of box C.

☒ Patent family members are listed in annex.

* Special categories of cited documents:

- *A* document defining the general state of the art which is not considered to be of particular relevance
- *E* earlier document but published on or after the International filing date
- *L* document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)
- *O* document referring to an oral disclosure, use, exhibition or other means
- *P* document published prior to the international filing date but later than the priority date claimed

- *T* later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention
- *X* document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone
- *Y* document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art
- *G* document member of the same patent family

Date of the actual completion of the international search

5 May 2004

Date of mailing of the international search report

26/05/2004

Name and mailing address of the ISA

European Patent Office, P.B. 5818 Patentlaan 2
NL - 2280 HV Rijswijk
Tel (+31-70) 340-2040, Tx 31 651 epo nl,
Fax (+31-70) 340-3016

Authorized officer

Lersch, W

INTERNATIONAL SEARCH REPORT

International Application No
PCT/IB2004/050131

C.(Continuation) DOCUMENTS CONSIDERED TO BE RELEVANT

Category *	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
A	J.H.BRITTAIN ET AL.: "Whole-Body Imaging with a Continuously Moving Patient Table" PROC.INTL.SOC.MAG.RESON.MED. 10, 2002, P. 2336, XP002279012 conference abstract -----	4,10

INTERNATIONAL SEARCH REPORT

Information on patent family members

International Application No
PCT/IB2004/050131

Patent document cited in search report		Publication date		Patent family member(s)	Publication date
US 2002021128	A1	21-02-2002	JP	2002010992 A	15-01-2002
US 2003004408	A1	02-01-2003	NONE		
EP 1398642	A	17-03-2004	US	2004051529 A1	18-03-2004
			EP	1398642 A2	17-03-2004
EP 1205760	A	15-05-2002	US	6445181 B1	03-09-2002
			EP	1205760 A2	15-05-2002